

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”


IS/ISO 11979-4 (2008): Ophthalmic Implants - Intraocular Lenses, Part 4: Labelling Information [MHD 5: Ophthalmic Instruments and Appliances]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



IS/ISO 11979-4 : 2008
(Superseding IS 14323 : 1996)

भारतीय मानक
नेत्र रोग संबंधी रोपण — इंद्राऑक्यूलर लेंस
भाग 4 लेबलिंग एवं सूचना

Indian Standard
OPHTHALMIC IMPLANTS — INTRAOCULAR LENSES
PART 4 LABELLING AND INFORMATION

ICS 11.040.70

© BIS 2011
BUREAU OF INDIAN STANDARDS
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

June 2011

Price Group 3

NATIONAL FOREWORD

This Indian Standard (Part 4) which is identical with ISO 11979-4 : 2006 'Ophthalmic implants — Intraocular lenses — Part 4: Labelling and information' issued by the International Organization for Standardization (ISO) was adopted by the Bureau of Indian Standards on the recommendation of the Ophthalmic Instruments and Appliances Sectional Committee and approval of the Medical Equipment and Hospital Planning Division Council.

This standard is published in ten parts. Other parts of this series are:

- Part 1 Vocabulary
- Part 2 Optical properties and test methods
- Part 3 Mechanical properties and test methods
- Part 5 Biocompatibility
- Part 6 Shelf-life and transport stability
- Part 7 Clinical investigations
- Part 8 Fundamental requirements
- Part 9 Multifocal intraocular lenses
- Part 10 Phakic intraocular lenses

The IS/ISO 11979 series will supersede IS 14323 : 1996 'Intraocular lenses — Specification' and after the publication of this series of standards, IS 14323 shall be treated as withdrawn.

This standard contains requirements and guidance for the labelling of intraocular lenses and the information supplied with them. Labelling requirements for medical devices are given in order to ensure correct and necessary information to the ophthalmic surgeon. Some additional information is required for intraocular lenses. This information concerns technical and optical data as well as information about materials used.

The text of ISO Standard has been approved as suitable for publication as an Indian Standard without deviations. Certain conventions are, however, not identical to those used in Indian Standards. Attention is particularly drawn to the following:

- a) Wherever the words 'International Standard' appear referring to this standard, they should be read as 'Indian Standard'.
- b) Comma (,) has been used as a decimal marker while in Indian Standards, the current practice is to use a point (.) as the decimal marker.

For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test, shall be rounded off in accordance with IS 2 : 1960 'Rules for rounding off numerical values (*revised*)'. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

Indian Standard

OPHTHALMIC IMPLANTS — INTRAOCULAR LENSES

PART 4 LABELLING AND INFORMATION

1 Scope

This part of ISO 11979 specifies the labelling requirements for intraocular lenses (IOLs) and the information to be provided within or on the packaging.

NOTE This part of ISO 11979 attempts to harmonize the recognized labelling requirements for IOLs throughout the world. However, there might be additional national requirements.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11979-1, *Ophthalmic implants — Intraocular lenses — Part 1: Vocabulary*

ISO 11979-9:2006, *Ophthalmic implants — Intraocular lenses — Part 9: Multifocal intraocular lenses*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

EN 980, *Symbols for use in the labelling of medical devices*

EN 1041, *Information supplied by the manufacturer with medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11979-1 apply.

4 Information provision

The general requirements for information provision provided with medical devices by manufacturers included in EN 1041 shall be taken into account.

5 Labelling

Table 1 lists minimal information that shall be included with the labelling of IOLs and whereabouts on the packaging it shall be given. Table 2 lists information that shall be given in applicable cases.

Table 1 — Information that shall be included with the packaging of IOLs

Item	Information	Primary package ^{a, b} and/or additional wrapping(s)	Storage container ^b	Comments
1	Name or trade name of the manufacturer ^c	X	X	The manufacturer's logotype may be added.
2	Address of the manufacturer and country of manufacture		X	
3	Trade name and/or model designation of the product	X	X	
4	Batch code or serial number	X	X	Symbols may be used.
5	The word "STERILE"	X	X ^d	Symbols may be used.
6	Method of sterilization		X	Symbols may be used.
7	The statement "Do not resterilize"		X	
8	The statement "For single use"		X	Symbols may be used.
9	Expiration date (year and month; format: YYYY MM)		X	Symbols may be used.
10	Dioptric power	X	X	The unit is m ⁻¹ ; in ocular optics often denoted by the special symbol D.
11	Overall diameter, in millimetres	X	X	May be indicated by the symbol Ø or in a drawing.
12	Diameter (minimum and maximum dimensions, if non-circular) of the body, in millimetres	X	X	May be indicated in a drawing.
13	Drawing depicting the configuration of the lens		X	
14	Statement about intended placement		X	For instance: anterior chamber; posterior chamber; in the bag; etc.
15	Information aiding the surgeon to calculate the dioptric power to implant		X	Currently, there is no standardized methodology available.
16	Information on the recommended storage conditions for the lens		X	To be included where specific storage conditions are required.

^a There are different systems in use with regard to primary container and additional wrapping(s). The information listed in the column is to be given on the appropriate component as to guarantee safe use and proper handling of the device.

^b In ISO 11607-1 and ISO 11607-2 "primary package" is now "sterile barrier system"; "storage container" is now "protective packaging".

^c See definition of "manufacturer" in ISO 11979-1.

^d If a statement about sterility is made on the storage container, it is to read "contains (number) sterile IOL(s)", if the storage container contains more than one IOL.

Table 2 — Information that shall be included with the packaging of IOLs in applicable cases

Item	Information	Primary package and/or additional wrapping(s)	Storage container	Comments
17	Additional descriptions the manufacturer wishes to provide		X	For instance: about optic shape, type and material; haptic type and material; UV-absorber; foldable; etc.
18	Statement “Custom made device”		X	
19	Statement “Exclusively for clinical investigation”		X	
20	For multifocal intraocular lenses: base power and add power(s)		X	

6 Package insert

The package insert, in the form of a leaflet or similar, shall be included in the storage container in such a way that it can be consulted without damage to the sterile packaging. It shall contain at least the following information.

- a) name or trade name and address of the manufacturer;
- b) detailed description of the lens including material(s) used;
- c) method of sterilization;
- d) conditions of storage and transport (if appropriate);
- e) instructions for the removal of the IOL from the primary container;
- f) instructions for use;

NOTE European Directive 93/42/EEC as amended by Directive 2007/47/EC requires date of issue or the latest revision of the instructions for use for medical devices sold in Europe.

- g) indication(s) of the circumstances under which the IOL can be used;
- h) contra-indication(s) of the circumstances under which the IOL should not be used;
- i) complication(s) that may occur;
- j) warning not to implant the IOL if the container which maintains sterility has been opened or damaged;
- k) warning not to re-use the IOL;
- l) warning not to re-sterilize the IOL;
- m) for multifocal and phakic intraocular lenses, a summary of the results of the clinical investigation, if any;
- n) for multifocal intraocular lenses, a graph of the MTF (modulation transfer function) through focus response performance of the intraocular lens in the model eye, using the conditions described in Annex A of ISO 11979-9:2006; informative text shall accompany the figure explaining that the MTF values in the graph describe the lens' optical performance in a standardized eye model at 50 c/mm as the focus is gradually shifted from that of a far object to increasingly near objects and that higher numbers indicate

better performance, and a graph of spectral transmittance through the lens in the range of 300 nm to 1 100 nm (see also ISO 11979-2);

- o) for phakic intraocular lenses, recommendations for periodic evaluations after implantation, based on risk analysis and/or any clinical investigation performed, and a restriction in the indications for use if necessitated by the anatomical clearance analysis and clinical evaluation;
- p) other appropriate warning(s).

7 Self-adhesive label

If supplied, a self-adhesive label shall contain at least the following information.

- a) name or trade name of the manufacturer;
- b) trade name and/or model designation of the product;
- c) batch code or serial number (symbols may be used);
- d) dioptric power;
- e) add power (for multifocal IOLs);
- f) overall diameter (may be indicated by the symbol $\varnothing$ or in a drawing);
- g) diameter of the body (minimum and maximum dimensions, if non-circular); this can be indicated in a drawing.

8 Use of symbols

Symbols can be used instead of text where appropriate. When symbols are used, ISO 15223-1 or EN 980, depending on region, apply.

Bibliography

- [1] ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*
- [2] ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act*, 1986 to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Director (Publications), BIS.

Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the latest issue of 'BIS Catalogue' and 'Standards: Monthly Additions'.

This Indian Standard has been developed from Doc No.: MHD 05 (0165).

Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002

Telephones: 2323 0131, 2323 3375, 2323 9402

Website: www.bis.org.in

Regional Offices:

Telephones

Central	: Manak Bhavan, 9 Bahadur Shah Zafar Marg NEW DELHI 110002	{ 2323 7617 2323 3841
Eastern	: 1/14 C.I.T. Scheme VII M, V.I.P. Road, Kankurgachi KOLKATA 700054	{ 2337 8499, 2337 8561 2337 8626, 2337 9120
Northern	: SCO 335-336, Sector 34-A, CHANDIGARH 160022	{ 260 3843 260 9285
Southern	: C.I.T. Campus, IV Cross Road, CHENNAI 600113	{ 2254 1216, 2254 1442 2254 2519, 2254 2315
Western	: Manakalaya, E9 MIDC, Marol, Andheri (East) MUMBAI 400093	{ 2832 9295, 2832 7858 2832 7891, 2832 7892

Branches : AHMEDABAD. BANGALORE. BHOPAL. BHUBANESHWAR. COIMBATORE. DEHRADUN. FARIDABAD. GHAZIABAD. GUWAHATI. HYDERABAD. JAIPUR. KANPUR. LUCKNOW. NAGPUR. PARWANOO. PATNA. PUNE. RAJKOT. THIRUVANANTHAPURAM. VISAKHAPATNAM.